

tiamo 2.5 light

Declaration of Conformity

1 Declaration of Conformity

Description

Comprehensive and user-friendly PC program for the freely programmable control of titration networks comprising Titrandos, Titrinos, sample changers, balances and other devices. Creation of methods, carrying out titrations, storage and recalculation of results. Dialogs in English and German.

- **6.6056.251 tiamo 2.5 light**
number and type of instruments restricted, no client/server support, no compliance with FDA 21 CFR Part 11

Quality system

The software mentioned above was developed in accordance with the requirements of the ISO 9001 quality system regarding the design, testing and servicing of Metrodata software.


Technical specifications

The technical specifications are documented in the installation manual.

Software validation

The software was validated with respect to functionality, analytical performance and accuracy of results. The software functions are documented in the online help.

Herisau, July 16, 2020



P. Hunziker

Vice President Engineering



Dr. M. Saba

Vice President Quality Management
and Regulatory Affairs