

942 Extension Module Vario

Declaration of conformity

2024-01-08 8.942.3004EN / v4

Metrohm AG lonenstrasse CH-9100 Herisau Switzerland +41 71 353 85 85 info@metrohm.com www.metrohm.com

1 EU declaration of conformity / UK declaration of conformity

This declaration attests the compliance of the instrument with the standard specifications for electrical instruments and accessories.

The responsibility for issuing this declaration of conformity lies solely with the manufacturer.

1.1 Product validity

This declaration is valid for the following products or product versions:

942 Extension Module Vario

The 942 Extension Module Vario is an extension module for the Professional IC Vario line of instruments. It permits the integration of additional functionality in a Professional IC Vario system.

1.2 Directives and regulations

1.2.1 EU directives

The object of the declaration described above is in conformity with the relevant European Union harmonization legislation:



2014/35/EU – Low Voltage Directive, LVD

Directive 2014/35/EU of the European Parliament and of the Council of February 26, 2014 on the harmonization of the laws of the Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits; Official Journal of the EU L96, 29.03.2014, p. 357-374

2014/30/EU – EMC Directive, EMC

Directive 2014/30/EU of the European Parliament and of the Council of February 26, 2014 on the harmonization of the laws of the Member States relating to electromagnetic compatibility; Official Journal of the EU L96, 29.03.2014, p. 79-106

Safety specifications

 2011/65/EU – Directive for certain hazardous substances, RoHS

Directive 2011/65/EU of the European Parliament and of the Council of June 8, 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment; Official Journal of the EU L174, 01.07.2011, p. 88-110

2015/863/EU – Amendment Appendix II RoHS
 Commission delegated directive (EU) 2015/863 of March 31, 2015
 amending Appendix II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substan

2012/19/EU – Disposal and recycling of electrical and electronic equipment, WEEE

ces; Official Journal of the EU L137, 04.06.2015, p. 10-12

Directive 2012/19/EU of the European Parliament and of the Council of July 4, 2012 on waste electrical and electronic equipment (WEEE); Official Journal of the EU L197, 24.07.2012, p. 38-71

1.2.2 UK regulations

The object of the declaration described above is in conformity with the relevant statutory requirements of the United Kingdom:



- S.I. 2016/1101 The Electrical Equipment (Safety) Regulations 2016
- S.I. 2016/1091 The Electromagnetic Compatibility Regulations 2016
- S.I. 2012/3032 The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

1.3 Safety specifications

This instrument fulfills the following safety requirements:

Design and type testing

• EN 61010-1: 2019

Safety requirements for electrical equipment for measurement, control, and laboratory use

EN 61010-2-081: 2020

Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

EN 60529: 2013 – degree of protection IP20
Degrees of protection provided by housings (IP Code)

- ISO 12100: 2010

General principles for design - Risk assessment and risk reduction

EN 61140: 2016 – protection class I

Protection against electric shock - Common requirements for installations and equipment

2 -----

Testing in Production

Every instrument is routine-tested according to EN/IEC 61010-1 Appendix F in the production division:

Check of the protective conductor connection and the insulation from power circuits.

1.4 Electromagnetic compatibility (EMC)

Design and type testing

• EN 61326-1: 2013

Electrical equipment for measurement, control and laboratory use - General EMC requirements

1.4.1 Emission

Standards fulfilled

• EN 55011 / CISPR 11: 2020

1.4.2 Immunity

Standards fulfilled

■ EN 61000-4-2: 2009

■ EN 61000-4-3: 2010

• EN 61000-4-4: 2012

■ EN 61000-4-5: 2017

■ EN 61000-4-6: 2014

■ EN 61000-4-8: 2010

■ EN 61000-4-11: 2020

EN 61000-4-14: 2009

EN 61000-4-28: 2009

1.5 Manufacturer

Metrohm AG, Ionenstrasse, CH-9100 Herisau/Switzerland

Metrohm AG holds the SQS certificate ISO 9001: 2015 for quality management systems with scope for the development, production, sale and service of instruments and accessories for titration, ion chromatography, spectroscopy and electrochemistry, including software solutions.

Herisau, January 8, 2024

P. Hunzil

Patrick Hunziker

Vice President, Head of Development Dr. Miriam Saba

Dr. U. Loka

Vice President, Head of Quality Management & Regulatory Affairs