

## 1 Purpose

This document describes the changes, improvements and fixed bugs in the following firmware version:

- 5.900.0044

in comparison with the firmware version:

- 5.900.0043

## 2 New functions

- None

## 3 Improvements

The following improvements were made:

- LAN connection: For saving on shared memory and for printing with the network printer, the SMB protocol version 2 and 3 is supported.
- The behavior of the login dialog when the password has expired and when the maximum number of login attempts has been reached was improved. The validity of the password is only checked when logging in. If the password has expired, the user can use the **Change password** command directly.
- The saved buffer series for the pH calibration have been updated for the following buffer types: Fisher, Honeywell Fluka, Beckmann, Radiometer Analytical, Merck CertiPUR 20 / Titrisol and Merck CertiPUR 25.

## 4 Fixed bugs

The following bugs were fixed:

- The user interface sometimes froze when printing a report on the network printer.
- If a Remote Box and a dosing device 800/805 were connected to the same MSB connector, the method run was sometimes interrupted during the titration.

