

Release Notes

EN

1 Purpose

This document describes new features, improvements as well as fixed bugs and problems in the following new firmware version:

- 5.916.0040
- in comparison with the previous version.
- 5.916.0033

2 New features

- A new version of the 916 Ti-Touch is available, which is equipped with a built-in magnetic stirrer (article number: 1.916.0020).
- 916 Ti-Touch now complies with FDA 21 CFR Part 11.
- 916 Ti-Touch can now perform STAT titration.
- 916 Ti-Touch supports the new 810 Sample Processor.
- Added buffer pH 1.679 by "Radiometer Analytical".
- The document 8.840.4523 Titrando Compliance Guide 840/PC Control 4.1 has been updated for the 900 Touch Control and now includes 915 KF Ti-Touch, 916 Ti-Touch, and 917 Coulometer. New article number: 8.900.4007 Compliance Guide Touch Control.

3 Improvements

- The following buffer series have been updated: NIST (all), Fluka (all series), Merck CertiPUR 20/Titrisol (pH 12.000), Merck CertiPUR (25 °C) (pH 4.00 and pH 10.00).

4 Fixed Bugs and Problems

- If in the dialog **Dosing unit / PREP parameters** the volume was set to **cylinder volume**, the PREP run did not work correctly.
- If a determination was canceled in the dialog **Electronic signature**, the determination did not end correctly.

Herisau, May 23rd, 2017

Dr. H. Lohar

M. Saba

Head of Quality Management