



Autosampler MMS 5000 (AJ)

2.136.0800

Autosampler MMS 5000 (AJ) von Analytik Jena zum Einsatz mit dem Metrohm Combustion IC für die vollautomatische Analyse von flüssigen und festen Proben. Um den modularen Multi-Matrix-Probengeber an den richtigen Proben typ anzupassen, muss entweder das Flüssigkeitskit (6.7303.000) oder das Feststoffkit (6.7302.000) verwendet werden.

Lieferumfang 2.136.0800

Qt.	Order no.	Beschreibung
1 PCS	1.136.0800	MMS 5000 Autosampler (AJ)

MMS 5000 Autosampler (AJ) from Analytik Jena for use with the Metrohm Combustion IC for fully automatic analysis of liquid and solid samples. In order to match the modular Multi-Matrix sampler with the correct sample type, either the liquid kit (6.7303.000) or the solid kit (6.7302.000) must be used.



Optionales Zubehör

Order no.	Beschreibung	
6.7301.150	Substrate made of quartz for liquid samples Substrate for liquid sample injections, 3 pcs.	
6.7302.000	Kit for solid samples for MMS 5000, 35 positions Kit for solid samples for MMS 5000 with rack for sample boats, grippers and 10 quartz glass boats.	
6.7302.100	Quartz boat 40 x 9 mm (AJ) Quartz glass boat for sample injection in Combustion Module, 10 pcs.	
6.7302.310	Sample rack 35 positions for solid samples For solids.	

6.7303.000 Kit for liquid samples for MMS 5000, 112 positions

Kit for liquid samples for MMS 5000 with rack for sample vessels, 100 μ L syringe with replaceable needle and 100 snap-ring bottles.




6.7303.100 Glass vial, 1.5 mL, 100 pieces

For use with the following sample changers: Liquid Autosampler CIC (TEI), MMS 5000 Autosampler (AJ) and 889 Sample Center.



6.7303.110 Septum for 67303100, 100 pieces

For use with the 67303100 Glass vial on the following sample changers: Liquid Autosampler CIC (TEI), MMS 5000 Autosampler (AJ), and 889 Sample Center.



6.7303.200 50 μ L syringe with replaceable needle

Replacement syringe for MMS 5000 liquid setup



6.7303.210 100 μ L syringe with replaceable needle

Replacement syringe for MMS 5000 liquid setup



6.7303.220 Needles for 50 μ L/100 μ L syringes, 3 pcs.

Replacement needle for syringe for MMS 5000 liquid setup



6.7303.310 Sample rack 112 positions for liquid samples

For liquid samples.

