



Viscotrode

6.0239.100

Kombinierte pH-Elektrode für pH-Messung/Titration in:

- viskosen Proben (z.B. Nährlösungen, Emulsionen)
- eiweiss- oder sulfidhaltige Medien

Das flexible Schliffdiaphragma ist gegen Verschmutzung unempfindlich und besonders leicht zu reinigen.

Referenzelektrolyt: $c(\text{KCl}) = 3 \text{ mol/L}$, Aufbewahrung in Aufbewahrungslösung.

Optionales Zubehör

Order no.	Beschreibung
-----------	--------------

6.2104.020 Electrode cable / 1 m / F

For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).



6.2104.030 Electrode Cable 2 m / F

For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).



6.2308.020 Electrolyte 3 mol/L KCl (250 mL)

Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)



6.2323.000 Storage solution

Storage solution for all combined pH glass electrodes with reference electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$



6.2325.000

pHit kit

Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



6.2325.100

Cleaning solution 3 x 50 mL

Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode.

The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.



This cleaning solution is also part of the pHHit kit.
