



909 UV Digester (230 V)

2.909.0014

Digestion instrument for UV photolysis of water samples with low to medium organic load. For sample preparation in trace element determination by means of voltammetry, ion chromatography and spectroscopy (AAS, ICP). Integrated instrument with operating unit and wet end. With air cooling and automatic control of digestion temperature and time. For 12 samples with a maximum of 12 mL sample volume each. Instrument for 220 - 240 V and 50 - 60 Hz.

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 2.909.0014

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

1.909.0014

909 UV Digester (230 V)

Digestion instrument for UV photolysis of water samples with low to medium organic load. For sample preparation in trace element determination by means of voltammetry, ion chromatography and spectroscopy (AAS, ICP). Integrated instrument with operating unit and wet end. With air cooling and automatic control of digestion temperature and time. For 12 samples with a maximum of 12 mL sample volume each. Instrument for 220 - 240 V and 50 - 60 Hz.




1 PCS

6.1110.010

Pt1000 temperature sensor for 909 UV Digester

Pt1000 temperature sensor (class B) made of glass with fixed cable (0.5 m) for connecting to a 909 UV Digester.



12 PCS

6.1446.100

PTFE stopper for 12 mL UV quartz sample vessel

For 6.2414.000 quartz sample vessel



1 PCS

6.1446.190

Thermometer stopper for Pt1000 temperature sensor

For 6.1110.010 Pt1000 temperature sensor and 6.2414.000 12 mL quartz sample vessel.



1 PCS

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



1 PCS

6.2041.240

Sample holder for 12 quartz sample vessels 12 mL

For 909 UV Digester.



1 PCS

6.2225.000

UV warning plate

For 909 UV Digester



12 PCS

6.2414.000

Quartz sample vessel 12 mL

For UV Digester



1 PCS

6.2621.100

3 mm hex key for IC Sample Processors



1 PCS

6.2745.100

UV protective shield

For 909 UV Digester.



1 PCS

6.2804.090

UV mercury vapor lamp

Medium pressure mercury vapor lamp for 909 UV Digester.



Optional accessories

Order no.	Description	
6.2042.050	Spacer for Pt1000 temperature sensor	