

Combined Ag ring electrode

6.00450.100

Combined silver ring electrode with fixed ground-joint diaphragm.

This electrode is suitable for precipitation titrations when the pH value varies (titrant silver nitrate)

This electrode is suitable for precipitation titrations when the pH value varies (titrant silver nitrate of, e.g.:

- Chloride, bromide, iodide
- Sulfides
- Hydrogen sulfide
- Mercaptans
- Cyanides

The fixed ground-joint diaphragm is insensitive to contamination and the permanently fused-in silver ring is resistant to more highly concentrated acids and saline solutions.

 $c(KNO_3) = 1 \text{ mol/L}$ is used as reference electrolyte and storage solution.



Depending on the application, we recommend using an Ag ring electrode with a coating (AgBr, AgCl or Ag_2S coating), which can be ordered accordingly.

Below, the accessories are grouped into Scope of delivery and Optional accessories. Please keep this printout at hand for ordering replacement material. These lists may be subject to change.

Scope of delivery 6.00450.100

Qt.	Order no.	Description	
1 PCS	6.1236.050	Sleeve with SGJ 14/12 mm	
	Sleeve with SGJ 1	4/12 mm, polyethylene.	
1 PCS	6.2008.040	Storage vessel	
	Together with 6.2 the electrode on	2008.050 storage vessel holder. it provides a support for 807 Dosing Units.	



Optional accessories

Order no.	Description		
6.02104.000	Electrode cable plug-in head G / plug P, 0.55 m		
	Cable for connecting an analog electrode (plug-in head G) to an analog measuring module from OMNIS.		
6.02104.010	Electrode cable plug-in head G / plug P, 1.5 m		
	Cable for connecting an analog electrode (plug-in head G) to an analog measuring module from OMNIS.		
6.2104.020	Electrode cable / 1 m / F		
	For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).		
6.2310.010	Electrolyte KNO3-1mol/L 250 mL		
	Electrolyte solution KNO ₃ 1 M (reference electrolyte for combined silver electrodes)		