



Drain channel for OMNIS Sample Robot M/L WSM

6.01804.460

The drain channel is mounted behind the workstations of the OMNIS Sample Robot WSM M/L and, in case of an error, collects any overflowing solution in order to channel it safely and contact-free through tubing into a waste canister.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.01804.460

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.01803.000

Silicone tubing 8/12 2.5 m

Silicone tubing for connecting to the drainage of the Pick&Place module or pump module in order to channel any solution overflows that may occur safely into the waste canister.



1 PCS

6.01808.010

Y tubing connector, PP, 3 x 7-8 mm

Connector for waste tubing



1 PCS

6.1602.115

Bottle attachment / S 40 / green

For direct dosing from reagent bottles with S 40 thread (Merck...).



1 PCS

6.1624.000

Adapter SGJ 14 for 6.1619.XXX Adsorber tube

Adapter SGJ 14. Part of 6.1619.010 for Dilution Sample Processor.



