

iAquatrode Plus with Pt1000

6.0277.300

Intelligent, combined pH electrode with integrated memory chip for storing sensor data and with Pt1000 temperature sensor for pH measurements/titrations in ion-deficient aqueous media (e.g., drinking water, process water). This electrode shows a very short response time in these samples.

The fixed ground-joint diaphragm is insensitive to contamination.

When c(KCI) = 3 mol/L is used as bridge electrolyte, storage in storage solution is recommended. The bridge electrolyte can be easily replaced with a chloride-free electrolyte (e.g., potassium nitrate $c(KNO_3) = 1 \text{ mol/L}$ (6.2310.010)). Storage in the used bridge electrolyte.

iTrodes can be connected to Titrando, Ti-Touch, or 913/914 meters.

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 6.0277.300

Qt. Order no. Description



Sleeve with SGJ 14/12 mm, polyethylene.



1 PCS

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Optional accessories

Order no.	Description
2.854.0010	854 iConnect — Electrode cable and measuring amplifier for intelligent "iTrodes" electrodes (cable length 150 cm).
6.2308.020	Electrolyte 3 mol/L KCI (250 mL) Electrolyte solution c(KCI) = 3 mol/L, (for Ag/AgCI reference systems)
6.2310.010	Electrolyte KNO3-1mol/L 250 mL Electrolyte solution KNO ₃ 1 M (reference electrolyte for combined silver electrodes)
6.2323.000	Storage solution Storage solution for all combined pH glass electrodes with reference electrolyte c(KCl) = 3 mol/L



Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



6.2325.100 Cleaning solution 3 x 50 mL

Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode.



The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.

This cleaning solution is also part of the pHit kit.