



IC equipment, connection Agilent ICP-MS

6.05330.400

IC equipment for the connection and bidirectional control of Metrohm IC and Agilent ICP-MS (IC-ICP/MS)

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.05330.400

Qt.	Order no.	Description
1 PCS	6.2141.410	Remote cable, IC - Agilent ICP-MS

Connecting cable for synchronization of Metrohm IC Systems (930 IC Flex and 940 Professional IC) and Agilent ICP-MS systems



1 PCS

6.2141.420

Connecting cable, Remote Box MSB

Connecting cable between connection box and Remote Box MSB



1 PCS

6.2141.430

Connecting cable, leak sensor

Connecting cable between connection box and leak sensor inlet of the IC



1 PCS

6.2149.310

Connection Box, remote control, ICP-MS

Connector for use as interface between Metrohm IC (Remote Box MSB) and Agilent ICP-MS in MassHunter 5.3 and higher.

