



## 5-ring conductivity measuring cell $c = 0.7 \text{ cm}^{-1}$ with Pt1000 (fixed cable)

6.0915.100

5-ring conductivity measuring cell with cell constant  $c = 0.7 \text{ cm}^{-1}$  (guide value), with integrated Pt1000 temperature sensor and with fixed cable (1.2 m) for connecting to an 856 Conductivity Module.

This sensor is suitable for measurements of medium conductivities ( $5 \mu\text{S/cm}$  to  $20 \text{ mS/cm}$ ), e.g., in:



- drinking water
- surface water
- wastewater

Below, the accessories are grouped into Scope of delivery and Optional accessories.





Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 6.0915.100

Qt.	Order no.	Description
1 PCS	6.1236.050	Sleeve with SGJ 14/12 mm Sleeve with SGJ 14/12 mm, polyethylene.
		
1 PCS	6.2008.040	Storage vessel Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.
		

# Optional accessories

Order no.	Description	
6.2301.060	KCl ion standard 250 mL	
	KCl ion standard ( $c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$ ), can also be used as standard solution for conductometry $12.87 \text{ mS/cm}$ ( $25^\circ\text{C}$ ).	
6.2324.010	Conductivity standard $100 \mu\text{S/cm}$ 250 mL	
	Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$ .	
6.2324.110	Conductivity standard $100 \mu\text{S/cm}$ , 5 x 30 mL	
	Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$ .	
6.2763.000	Flow-through cell for the 6.0915.100/6.0920.100	
	Flow-through cell for the 6.0915.100/6.0920.100 5-ring conductivity measuring cell	