



# Metrosep A PCC 2 VHC/4.0

6.1006.350

For anion preconcentration and matrix elimination. The very high capacity is required primarily when matrix effects might cause an overloading of the preconcentration column or when samples with high ionic strength are to be analyzed.

Below, the accessories are grouped into Scope of delivery and Optional accessories. Please keep this printout at hand for ordering replacement material. These lists may be subject to change.

## Scope of delivery 6.1006.350

Qt.	Order no.	Description
2 PCS	6.2744.060	Threaded stopper
		For UNF 10/32. Stopper for IC, e.g. for the sealing of columns.

