



Metrosep A Supp 16 - 150/2.0

6.1031.220

The Microbore version of the Metrosep A Supp 16 - 150 is well-suited to medium-capacity separation problems. Eluent consumption is drastically reduced as a result of the smaller inner diameter of this column type and the correspondingly lower flows. As a result of the lower flows, the dwell time of the anions in the detector, and thus also the peak areas with identical sample quantities, are increased. The Microbore separation columns are used together with the MSM-LC. The 2 mm Metrosep A Supp 16 separation columns are packed with the same material as the corresponding 4 mm separation columns. The medium version of this column type is used for universal applications.

The column is very well-suited to applications that have a high ionic load but that do not require the highest resolution. With its low eluent flow, this column is particularly suitable for IC-MS coupling.

Below, the accessories are grouped into Scope of delivery and Optional accessories.

Please keep this printout at hand for ordering replacement material.

These lists may be subject to change.

Scope of delivery 6.1031.220

Qt.	Order no.	Description
-----	-----------	-------------

2 PCS



6.2744.060

Threaded stopper

For UNF 10/32. Stopper for IC, e.g. for the sealing of columns.



Optional accessories

Order no.	Description	
6.1031.600	Metrosep A Supp 16 Guard/2.0	
	The Metrosep A Supp 16 Guard/2.0 reliably protects the Metrosep A Supp 16 analytical separation columns with 2 mm inner diameter against contamination. Thanks to the "On Column Guard System", the guard column is very easy to handle. The guard column screws easily onto the analytical column. No tools are required.	
6.1031.610	Metrosep A Supp 16 S-Guard/2.0	
	The Metrosep A Supp 16 S-Guard/2.0 reliably protects the Metrosep A Supp 16 analytical separation columns with 2 mm inner diameter against contamination. The guard column is connected to the separation column simply using capillary connections. No tools are required.	