



# Sample rack 211 x 11 mL

6.2041.770

Multi-row sample rack for 211 samples in 11 mL sample vessels (6.2743.050)

Below, the accessories are grouped into Scope of delivery and Optional accessories.  
Please keep this printout at hand for ordering replacement material.  
These lists may be subject to change.

## Optional accessories

Order no.	Description
6.2743.040	Vial PP, 2.5 mL, 2000 pieces
	<p>2.5 mL. 2000 pieces. For IC Sample Processors.</p> <p>The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean-room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.</p>



6.2743.047 Vial PP, 2.5 mL, 200 pieces

2.5 mL. 200 pieces. For IC Sample Processors.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



---

6.2743.050 Sample vessel 11 mL

2000 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



---

6.2743.057 Sample vessel 11 mL

200 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



---

6.2743.070 Stopper with perforation

2000 pieces. For IC Sample Processors



6.2743.077 Stopper with perforation  
200 pieces. For IC Sample Processors



---

6.2841.120 Liquid Handling Station, left

Combined rinsing and dilution station for sample changers with a Swing Head. To be assembled to the left of the tower.



---

6.2841.130 Liquid Handling Station, right

Combined rinsing and dilution station for sample changers with two towers and a Swing Head. To be assembled to the right of the tower.



---

6.5330.130 IC equipment: Liquid Handling Station, left

Accessory set for assembling the Liquid Handling Station on the Professional Sample Processor. The rinsing water supply can be accomplished by means of a peristaltic pump or a Dosino.

