



Célula de medida de la conductividad con $c = 0,1 \text{ cm}^{-1}$ con Pt1000 (cable fijo, 0,65 m)

6.00921.040

Célula de medida de la conductividad de acero fino con constante de célula $c = 0,1 \text{ cm}^{-1}$ (valor guía), con sensor de temperatura Pt1000 integrado y cable fijo (0,65 m) para su conexión al OMNIS Measuring Module Conductivity.

Este sensor es apto para medidas de conductividades muy bajas (desde 0 hasta $300 \mu\text{S}/\text{cm}$), por ejemplo, en el agua desionizada o para medidas según la USP <645>.

Partes/accesorios 6.00921.040

Qt.	Order no.	Descripción
1 PCS	6.1236.050	Sleeve with SGJ 14/12 mm
		Sleeve with SGJ 14/12 mm, polyethylene.



Accesorios opcionales

Order no.	Descripción
-----------	-------------

6.02101.020 Measuring module conductivity

Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.



6.2324.010 Conductivity standard 100 µS/cm 250 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2324.110 Conductivity standard 100 µS/cm, 5 x 30 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.

