



# Electrodo de tensioactivos NIO

6.0507.010

Este electrodo de tensioactivos debe utilizarse en combinación con un electrodo de referencia y es adecuado para, por ejemplo:

- Titulación de tensioactivos no iónicos en matrices acuosas
- Titulación de tensioactivos a base de aductos de polioxietileno
- Titulación de principios activos farmacéuticos con tetrafenilborato de sodio

## Partes/accesorios 6.0507.010

| Qt.   | Order no.  | Descripción                          |
|-------|------------|--------------------------------------|
| 1 PCS | 6.1236.020 | Sleeve with SGJ 14/12 mm             |
|       |            | Sleeve with SGJ 14/12 mm and O-ring. |



Together with 6.2008.050 storage vessel holder, it provides a support for the electrode on 807 Dosing Units.



## Accesorios opcionales

| Order no.  | Descripción                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                       |
|------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------|
| 6.0726.100 | <p>Ag/AgCl reference electrode (length 12.5 cm)</p> <p>Silver/silver chloride reference electrode with Double-Junction System, installation length 10 cm.</p> <p>The standard ground-joint 14/15 enables easy assembly, and the flexible ground-joint diaphragm, which is insensitive to contamination, can be replaced at any time. The reference electrolyte and the bridge electrolyte can be selected freely according to use, and are easy to replace. This sensor is supplied without electrolyte filling.</p>                                                                                                                                                               |    |
| 6.0750.100 | <p>LL ISE reference electrode</p> <p>Silver / silver chloride reference electrode with double junction system. This reference electrode is well suited for:</p> <ul style="list-style-type: none"><li>• automated applications</li><li>• ion measurements</li><li>• surfactant titrations</li></ul> <p>The ground-joint diaphragm, which is insensitive to contamination, offers a constant and reproducible electrolyte outflow. Additionally, the reference electrolyte is gelified for even better signal stability. The sensor is delivered with <math>c(\text{KCl}) = 3 \text{ mol/L}</math> as bridge electrolyte, which can be freely selected and exchanged as needed.</p> |   |
| 6.2104.020 | <p>Electrode cable / 1 m / F</p> <p>For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |  |