



856 Conductivity Module

2.856.0010

Modulo di misura della conducibilità come estensione di un sistema Titrande esistente oppure come strumento «stand alone» in combinazione con un 900 Touch Control. Con l'856 Conductivity Module possono essere determinate sia conducibilità, sia temperatura, nonché TDS e salinità. Supporta le celle di misura della conducibilità di ultima tecnologia, le celle di misura a 5 anelli.

Il Conductivity Module è dotato di 2 interfacce USB per il collegamento di stampanti, lettori di codice a barre o campionatori e di 4 interfacce MSB per agitatori o Dosino.

Utilizzo con OMNIS Software, software tiamo o Touch Control. Conforme ai requisiti GMP/GLP e FDA, nonché 21 CFR Parte 11, se necessario.

Parti incluse 2.856.0010

Qt.	Order no.	Descrizione
-----	-----------	-------------

Conductivity measuring module as supplement to an existing Titrand system or "stand-alone" in combination with a 900 Touch Control. With the 856 Conductivity Module, not only conductivity and temperature can be determined, but also TDS and salinity. It supports state-of-the-art conductivity measuring cells, i.e. 5-ring measuring cells.




Thanks to the galvanically isolated measuring input, pH value and conductivity can be measured in the same beaker without interference.



The Conductivity Module is equipped with two USB interfaces for connecting printers, barcode readers or sample changers and four MSB interfaces for stirrers or Dosinos.

Both in conjunction with the 900 Touch Control as stand-alone instrument and also integrated in *tiamo*™ full (from 2.0), it is in compliance with GLP and FDA 21 CFR part 11 requirements.

Accessori opzionali

Order no.	Descrizione	
2.801.0010	801 Stirrer	
	Magnetic stirrer without stand for supplementing the Titrino plus, Dosimat plus, Titrandos, Sample Processors, 805 Dosimats and 780/781 pH Meters. With permanently attached cable for MSB (Metrohm Serial Bus).	
6.0915.100	5-ring conductivity measuring cell $c = 0.7 \text{ cm}^{-1}$ with Pt1000 (fixed cable)	
	5-ring conductivity measuring cell with cell constant $c = 0.7 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (1.2 m) for connecting to an 856 Conductivity Module. This sensor is suitable for measurements of medium conductivities ($5 \mu\text{S/cm}$ to 20 mS/cm), e.g., in: <ul style="list-style-type: none">• drinking water• surface water• wastewater	
6.0915.130	5-ring conductivity measuring cell $c = 1.0 \text{ cm}^{-1}$ with Pt1000 (fixed cable)	
	5-ring conductivity measuring cell with cell constant $c = 1.0 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (1.2 m) for connecting to an 856 Conductivity Module in combination with a sample changer. This sensor is suitable for automated measurements of medium conductivities ($5 \mu\text{S/cm}$ to 100 mS/cm), e.g., in: <ul style="list-style-type: none">• drinking water• surface water• wastewater	

Conductivity measuring cell made of stainless steel with cell constant $c = 0.1 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and fixed cable (1.2 m) for connecting to an 856 Conductivity Module.

This sensor is suitable for measurements of low conductivities ($0 \text{ }\mu\text{S/cm}$ to $300 \text{ }\mu\text{S/cm}$) in e.g., deion. water or for measurements in accordance with USP <645>.



5-ring conductivity measuring cell with cell constant $c = 0.7 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and fixed cable (2.0 m) for connecting to an 856 Conductivity Module in combination with a sample changer.

This sensor is suitable for automated measurements of medium conductivities ($5 \text{ }\mu\text{S/cm}$ to 20 mS/cm), e.g., in:

- drinking water
- surface water
- wastewater



5-ring conductivity measuring cell with cell constant $c = 1.0 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and fixed cable (2.0 m) for connecting to an 856 Conductivity Module in combination with a sample changer.

This sensor is suitable for automated measurements of medium conductivities ($5 \text{ }\mu\text{S/cm}$ to 100 mS/cm), e.g., in:

- drinking water
- surface water
- wastewater



6.2065.000 Stacking frame for 846 Dosing Interface, 856 Conductivity Module, 867 pH Module

For fixing the Reagent Organizer on top of the Dosing Interface



6.2103.160 Adapter 4 x socket B – plug N

Adapter box for the connection of classical Metrohm Conductivity measuring cells with 4 banana plugs to the 856 Conductivity Module.



6.2151.000 Cable USB A – mini-DIN 8-pin

Controller cable

