



dAquatrode Plus con Pt1000

6.00202.300

Elettrodo per pH combinato digitale per OMNIS con rilevatore di temperatura Pt1000 integrato per misure del pH/titolazione in mezzi acquosi poveri di ioni (ad es. acqua potabile, acqua di processo). Questo elettrodo presenta un tempo di risposta particolarmente rapido anche in questi campioni.

Il diaframma fisso a smeriglio è insensibile alle impurità.

In caso di utilizzo di $c(\text{KCl}) = 3 \text{ mol/L}$ come elettrolita intermedio si raccomanda la conservazione in una soluzione di conservazione.

L'elettrolita intermedio può essere sostituito con un elettrolita privo di cloruro (ad es. nitrato di potassio $c(\text{KNO}_3) = 1 \text{ mol/L}$ (6.2310.010)). Conservazione nell'elettrolita utilizzato.

Gli elettrodi dTrode possono essere utilizzati con i titolatori OMNIS.

Parti incluse 6.00202.300

Qt.	Order no.	Descrizione
-----	-----------	-------------

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



1 PCS

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Accessori opzionali

Order no.	Descrizione	
6.02104.300	Electrode cable plug-in head Q / plug P, 0.55 m Cable for connecting a dTrobe (plug-in head Q) to the digital measuring module from OMNIS.	
6.02104.310	Electrode cable plug-in head Q / plug P, 1.5 m Cable for connecting a dTrobe (plug-in head Q) to the digital measuring module from OMNIS.	
6.2308.020	Electrolyte 3 mol/L KCl (250 mL) Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)	
6.2310.010	Electrolyte KNO_3 -1 mol/L 250 mL Electrolyte solution KNO_3 1 M (reference electrolyte for combined silver electrodes)	

6.2323.000

Storage solution

Storage solution for all combined pH glass electrodes with reference electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$



6.2325.000

pHit kit

Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



6.2325.100

Cleaning solution 3 x 50 mL

Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode.

The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.



This cleaning solution is also part of the pHit kit.