



Beaker per campioni in vetro trasparente da 75 mL, 1 pezzo

6.1432.210

Beaker per campioni in vetro trasparente da 75 mL, adatto ai rack dei campioni 6.2041.340, 6.2041.350, 6.2041.800 e 6.2041.900, 1 pezzo

Accessori opzionali

Order no.	Descrizione
6.1458.020	Titration head /4 x M10 connectors With 4 M10 screw nipples, including stoppers.



6.1458.030 KF titration head / 4 x M10 connectors

With 4 M10 screw nipples, including stoppers.



6.1458.070 Titration head / 2 x SGJ 14

Titration head for 75 mL beakers. With 2 openings SGJ 14, 2 openings 6.4 mm, and 3 pre-mounted dosing tips. Suitable for rinsing and aspiration equipment.



6.1909.030 Stirring propeller / 95 mm

Propeller stirrer with PP shaft, PVDF tip. Length from the SGJ: 104 mm. For KF applications with 722, 802 Rod Stirrer



6.2037.040 Foil holder

Foil holder for 75 mL sample beaker for Karl Fischer for KF Sample Changer



6.2041.340 Sample rack, 24 x 75 mL, (PVC)

Sample rack for 810 Sample Processor, 814 USB Sample Processor, 855 Robotic Titrosampler "Light", suitable for 24 sample beakers. The following sample beakers can be used: 6.1432.210, 6.1459.400.

Plastic: Polyvinyl chloride (PVC)



6.2041.350 Sample rack, 48 x 75 mL (PVC)

Sample rack for 815 Robotic USB Sample Processor XL, 855 Robotic Titrosampler "Basic", suitable for 48 sample beakers. The following sample beakers can be used: 6.1432.210, 6.1459.400.

Plastic: Polyvinyl chloride (PVC)



6.2041.800 Sample rack, 100 x 75 mL, (PVC)

Sample rack for 815 Robotic USB Sample Processor XL, 855 Robotic Titrosampler "Basic" suitable for 100 sample beakers. The following sample beakers can be used: 6.1432.210, 6.1459.400.

Plastic: Polyvinyl chloride (PVC)



6.2041.900 Sample rack insert, 54 x 75 mL, for water bath, (Al)

Sample rack insert for 815 Robotic USB Sample Processor XL, 855 Robotic Titrosampler "Basic", suitable for 54 sample beakers. The following sample beakers can be used: 6.1432.210, 6.1459.400.

Material: Aluminum (Al)



6.2820.000

Al foil

1000 circular sheets of 80 mm diameter. Thickness: 0.010 mm. Used together with 6.2037.030 Foil holder for 6.1432.210 Sample beaker. For 674 KF Sample Changer


