



# OMNIS Oven Module 6 mL

2.1017.1000

OMNIS Sample Robot Oven 6mL

**2.1017.1000**

Qt.	Order no.	
1 PCS	6.01807.010	Adsorber cartridge for OMNIS Oven Module

Adsorber cartridge suitable for the OMNIS Oven Module




1 PCS

**6.1446.230**

**Heating tubing stopper**

Heating tube stopper, for 1 titration cell (coulometry)



---

1 PCS

**6.1448.080**

**Septum, 7.5 mm, 5 pieces**

Replacement septums for the conditioning tube on an OMNIS Oven Module



---

1 PCS

**6.1805.470**

**FEP tubing / M6 / 44 cm**

With light and kink protection



---

1 PCS

**6.1808.040**

**Thread adapter M6 outer / M8 inner**

Outer thread M6, inner thread M8.



1 PCS

**6.1808.050**

**Tubing adapter olive / M8 outer**

1 M8 outer thread and 1 tubing olive. E.g. for thermostat jacket of exchange units and stability measuring instruments.



---

1 PCS

**6.1830.050**

**OMNIS heating tubing**

Heating tubing with M6 thread for OMNIS Sample Robot Oven



---

1 PCS

**6.2049.060**

**11–55 mm distance holder set**

Distance holder set for the OMNIS Sample Robot Oven to adjust the immersion depth of the needle in the sample vial.



---

1 PCS

**6.2621.140**

**Hexagon key 2.5 mm**



1 PCS

**6.2724.010**

**Dust filter**

Dust filter for Rancimats and Karl Fischer ovens.



1 PCS

**6.2739.000**

**Wrench**

For tightening connectors



1 PCS

**6.2811.000**

**Molecular sieve**

Molecular sieve. Bottle containing 250 g. Pore size: 0.3 nm. Without moisture indicator. For Rancimats and Karl Fischer instruments.



1 PCS

**6.2816.070**

**Septum piercing needle**

Used with Karl Fischer ovens.



1 PCS

**6.2816.080**

**Outlet needle**

Used with KF Thermoprep and Oven Sample Processor.



---

1 PCS

**6.2821.090**

**Aspiration filter**

Pore size: 27 µm. Set of 5 pieces. For aspiration tubing (6.1834.000) and filter tubes (6.1821.040) and (6.1821.050).

