



863 Compact IC Autosampler

2.863.0010

863 IC

2.863.0010

Qt.	Order no.	
1 PCS	1.863.0010	863 Compact IC Autosampler

The 863 Compact IC Autosampler is the ideal help for routine analysis. It enables automation of the full range of Metrohm IC instruments.



2 PCS

6.1826.360

Pump tubing LFL (white/white), 3 stoppers

For Sample Processors



1 PCS

6.1831.050

PEEK capillary 0.5 mm i.d., 40 cm

For inline dialysis and for sample changers. For IC Dialysis Unit, IC Sample Processor, Compact IC Autosampler, IC Dialysis Sample Processor, IC Liquid Handling Dialysis Unit, Compact VA Autosampler



1 PCS

6.1831.060

PEEK capillary 0.5 mm i.d., 1 m

For inline dialysis and for Sample Changers. For IC Dialysis Unit, IC Sample Processor, Compact IC Autosampler, IC Filtration Sample Processor, IC Dialysis Sample Processor, IC Liquid Handling Dialysis Unit



1 PCS

6.2151.100

Adapter USB MINI (OTG) - USB A

For connecting USB instruments.



1 PCS

6.2151.110

Metrohm USB Mini B cable (OTG) - USB A, 1.8 m

For connecting USB instruments.



1 PCS

6.2621.090

1/2 in. wrench

For 1/2 in. screws. For IC Sample Processors



1 PCS

6.2621.130

Hexagon key 2 mm

2 mm.



1 PCS

6.2621.140

Hexagon key 2.5 mm



1 PCS

6.2739.000

Wrench

For tightening connectors



1 PCS

6.2743.057

Sample vessel 11 mL

200 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



1 PCS

6.2743.077

Stopper with perforation

200 pieces. For IC Sample Processors



1 PCS

6.2744.034

Coupling olive/UNF 10/32, 2x

Connection between pressure screw and pump tubing. 2 pieces. For IC instruments with peristaltic pump.



1 PCS

6.2744.070

Pressure screw short

Short version. With UNF 10/32 connection. 5 pieces. For the connection of PEEK capillaries




1 PCS

6.2744.160

Nozzle/UNF 10/32 coupling with security device

Security device for connecting pump tubing to a nozzle.



1 PCS

6.2833.030

**Needle holder 1/8 in. for 863
Compact Autosampler**

Needle holder 1/8 in. for 863 Compact Autosampler



1 PCS

6.2846.010

**Sample needle made of zirconium oxide with
PEEK tip**

Sample needle for Sample Processors and Autosampler in ion chromatography. The PEEK tip guarantees secure piercing of the vial covering.



Order no.

6.1826.310 Pump tubing LFL (orange/green), 3-stopper

Pump tubing for suppressor solutions and for acceptor solutions for inline dialysis and for inline ultrafiltration.



6.1826.320 Pump tubing LFL (orange/yellow) , 3-stopper

For suppressor solutions and for acceptor solutions for inline dialysis and for inline ultrafiltration.



6.1826.330 Pump tubing LFL (orange/white), 3 stoppers

For all IC instruments with peristaltic pump.



6.1826.340 Pump tubing LFL (black/black), 3 stoppers

For the sample solution in inline dialysis.



6.1826.380

Pump tubing LFL (gray/gray), 3 stoppers

For Inline Dilution



6.1826.390

Pump tubing LFL (yellow/yellow), 3 stoppers

For the sample solution in Inline Ultrafiltration.



6.2141.230

**Remote cable Compact Sample Changer/Compact Autosampler -
Titrino plus/Pump Station**

Cable for the remote connection between Compact Sample Changer
/Compact Autosampler and Titrino plus and/or Pump Station.



6.2141.270

**Cable Compact IC Autosampler - Remote Connection on IC
Instruments**

Enables remote control of the Compact IC Autosampler from IC
Instruments



6.2151.020

Cable USB A - USB B 1.8 m

USB connecting cable



6.2629.000

Screw nut to needle holder IC

On Sample Processors



6.2723.310

Protective cover for Titrino plus

Protective cover for display and keypad for all instruments of the Titrino plus family.

Suitable for: 848 Titrino plus, 862 Compact Titrosampler, 863 Compact Autosampler, 865 Dosimat plus, 869 Compact Sample Changer, 870 KF Titrino plus, 876 Dosimat plus, 877 Titrino plus, 899 Coulometer.



6.2743.040

Vial PP, 2.5 mL, 2000 pieces

2.5 mL. 2000 pieces. For IC Sample Processors.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean-room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



6.2743.047

Vial PP, 2.5 mL, 200 pieces

2.5 mL. 200 pieces. For IC Sample Processors.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



6.2743.050

Sample vessel 11 mL

2000 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.

