



c = 0.5 cm-1Pt1000 (0.65 m)

6.00923.080

c = 0.5 cm⁻¹ () 4Pt1000OMNIS Measuring Module Conductivity (0.65 m)

PEEK (15 µS/cm 250 mS/cm) :

-
-
-

6.00923.080

Qt.

Order no.

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



Order no.**6.02101.020 Measuring module conductivity**

Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.

**6.2301.060 KCl ion standard 250 mL**

KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25°C).

**6.2324.010 Conductivity standard $100 \mu\text{S/cm}$ 250 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.

**6.2324.110 Conductivity standard $100 \mu\text{S/cm}$, 5 x 30 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.

