



iEcotrode Plus

6.0280.300

pH/

: c(KCl) = 3 mol / L

iTrodeTitrandoTi-Touch913/914

6.0280.300

Qt.		Order no.
1 PCS		6.1236.050 Sleeve with SGJ 14/12 mm
		Sleeve with SGJ 14/12 mm, polyethylene.



Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Order no.

2.854.0010

854 iConnect

The 854 iConnect – Electrode cable and measuring amplifier for intelligent "iTrodes" electrodes (cable length 150 cm).



6.2308.020

Electrolyte 3 mol/L KCl (250 mL)

Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)



6.2323.000

Storage solution

Storage solution for all combined pH glass electrodes with reference electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$



6.2325.000

pHit kit

Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode. The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.

This cleaning solution is also part of the pHit kit.

