



Titration Head, 3 x SGJ 14

6.1458.040

Titration head for 120 mL beakers. With 3 openings SGJ 14 and 4 openings 6.4 mm, 2 dosing tips included. Suitable for rinsing and aspiration equipment.

6.1458.040

Qt.	Order no.	
2 PCS	6.1808.000	Connection bushing M6



2 PCS

6.2726.090

Antidiffusion valve

For antidiffusion buret tips 6.1541.XXX, 6.1543.130, 6.1543.140,
6.1543.150, 6.1543.160, 6.1543.200



3 PCS

6.2740.020

Spray nozzle

For the fine-spraying of the rinsing solution.



Order no.

6.1829.050 Replacement tubings for retracted tubings in titration heads and robotic arms

Three replacement tubings with tapering tip.



6.1909.050 Stirring propeller / 102 mm

Propeller for 802 Stirrer when this is used with sample racks for 120 mL beakers (6.1459.300)



6.2041.470 Sample rack, 22 x 120 mL, (PP)

Sample rack for 810 Sample Changer, 814 USB Sample Processor, 855 Robotic Titrosampler "Light", suitable for 22 sample beakers. The following sample beakers can be used: 6.1459.300, 6.1459.307.

Plastic: Polypropylene (PP)



6.2841.010 External rinsing station for robotic arm with 6.1458.040 titration head

The external rinsing station can be fitted on the side next to the Robotic Sample Processor tower, making it possible for external cleaning of the titration equipment to be carried out during additional actions on the sample rack.

