



Porotrode

6.0235.200

Combined pH electrode for pH measurement / titration of:

- protein-containing samples (foodstuffs, biological samples)
- heavily contaminated samples
- viscous samples

The specially developed capillary diaphragm and the reference electrolyte porolyte (6.2318.000) enable an optimum performance in samples containing protein. Storage in storage solution.

Phạm vi giao hàng 6.0235.200

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



1 PCS




6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Phụ kiện tùy chọn

Order no.	Description	
6.2104.020	Electrode cable / 1 m / F For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).	
6.2104.030	Electrode Cable 2 m / F For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).	
6.2318.010	Porolyte 50 mL Reference electrolyte porolyte for electrodes 6.0235.200 Porotrode and 6.00235.600 Porotrode with Pt1000. Replaces 6.2318.000 Porolyte 50 mL.	
6.2325.100	Cleaning solution 3 x 50 mL Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode. The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life. This cleaning solution is also part of the pHit kit.	