



Separate Ag-rod electrode with Ag₂S coating

6.0331.010S

This metal electrode has to be used in combination with a reference electrode ($c(\text{KNO}_3) = 1 \text{ mol/L}$ as reference electrolyte) and is well suited for precipitation titrations (titrant silver nitrate). The robust shaft made of polypropylene and the metal rod can be replaced separately. The silver rod is coated with sulfide (Ag_2S) for higher sensitivity.

Phụ kiện tùy chọn

| Order no. | Description |
|-----------|-------------|
|-----------|-------------|

6.0729.100

Ag/AgCl reference electrode (length 12.5 cm)

Silver/silver chloride reference electrode with Double-Junction System and separate shielding, installation length 10 cm.

This reference electrode is well-suited for:

- Nonaqueous samples
- Differential potentiometry

The standard ground-joint 14/15 enables easy assembly, and the flexible ground-joint diaphragm, which is insensitive to contamination, can be replaced at any time. The reference electrolyte and the bridge electrolyte can be selected freely according to use, and are easy to replace. The sensor is supplied without electrolyte filling.



6.1241.030

Electrode shaft

Electrode shaft for separate metal pin electrodes



6.2106.020

Strand / 1 m / 2 x B

Electrode cable with 2 x plug B (4 mm) for connecting electrodes with plug-in head B



6.2310.010

Electrolyte KNO₃-1 mol/L 250 mL

Electrolyte solution KNO₃ 1 M (reference electrolyte for combined silver electrodes)

