



919 IC Autosampler plus for VA

2.919.0130

919 IC Autosampler plus for VA Professional-VA **viva viva** 28 10 mL

919 IC Autosampler plus for VA

Scope of delivery 2.919.0130

| Qt. | Order no. | Description |
|-------|------------|-------------------------|
| 1 PCS | 1.919.0020 | 919 IC Autosampler plus |

The 919 IC Autosampler plus fulfills the requirements of laboratories with medium sample numbers. It enables automation of the full range of Metrohm IC instruments.



1 PCS

6.1453.250

Sample beaker, plastic (PP), 250 mL, 1 piece

Sample beaker, plastic (polypropylene, PP), 250 mL, suitable for sample racks: 6.2041.310, 6.2041.400, 6.2041.430, 6.2041.510, 6.2041.820, 6.2068.020, and 6.2068.030, 1 piece



1 PCS

6.1803.160

PTFE capillary, 0.75 mm ID / 3 m

PTFE capillary for use with the 920 Absorber Module.



1 PCS

6.1815.010

Spiral band 0.5 m

For holding together different cables or tubing.




2 PCS

6.1826.020

PVC pump tubing (blue/blue)

For online IC instruments and automation in voltammetry.



1 PCS

6.1831.020

PEEK capillary 0.75 mm i.d., 22 cm

For all IC components.



1 PCS

6.2041.510

Sample rack, 56 x 11 mL and 1 x 250 mL (PVC)

Sample rack for 919 Autosampler plus, 858 Professional Sample Processor suitable for 56 sample vessels and 1 rinsing bottle. The following sample vessels and rinsing bottle can be used: 100 mL digestion tubes (CSB).

Plastic: Polyvinyl chloride (PVC)



1 PCS

6.2151.000

Cable USB A – mini-DIN 8-pin

Controller cable



1 PCS

6.2621.080

Capillary cutter

Cutting tool for simple cutting of plastic capillaries, 1/16" and 1/8", including 5 replacement blades.



1 PCS

6.2621.090

1/2 in. wrench

For 1/2 in. screws. For IC Sample Processors



1 PCS

6.2621.100

3 mm hex key for IC Sample Processors



1 PCS

6.2621.130

Hexagon key 2 mm

2 mm.



1 PCS

6.2739.000

Wrench

For tightening connectors



1 PCS

6.2743.057

Sample vessel 11 mL

200 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



2 PCS

6.2744.010

Pressure screw 5x

With UNF 10/32 connection. For the connection of PEEK capillaries




2 PCS

6.2744.160

Nozzle/UNF 10/32 coupling with security device

Security device for connecting pump tubing to a nozzle.



1 PCS

6.2833.030

**Needle holder 1/8 in. for 863
Compact Autosampler**

Needle holder 1/8 in. for 863 Compact Autosampler



1 PCS

6.2846.000

Sample needle made of zirconium oxide

Sample needle for Sample Processors and Autosamplers in voltammetry.

