



c = 0.5 cm⁻¹ Pt1000 2.0 m

6.00924.080

4 c = 0.5 cm⁻¹ () Pt1000 (2.0 m)
/15 µS/cm 250 mS/cm

-
-
-

Scope of delivery 6.00924.080

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



Optional accessories

Order no.	Description
-----------	-------------

6.02101.020 **Measuring module conductivity**

Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.



6.2301.060 **KCl ion standard 250 mL**

KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25°C).



6.2324.010 **Conductivity standard $100 \mu\text{S/cm}$ 250 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.



6.2324.110 **Conductivity standard $100 \mu\text{S/cm}$, 5 x 30 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.

