



# Unitrode

6.0259.100

pH pH

- pH
- 

cKCl = 3 mol / L

T>80°C Idrolyt

## Optional accessories

Order no.	Description
-----------	-------------

**6.2104.020**

**Electrode cable / 1 m / F**

For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).



---

**6.2104.030**

**Electrode Cable 2 m / F**

For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).



---

**6.2308.020**

**Electrolyte 3 mol/L KCl (250 mL)**

Electrolyte solution  $c(\text{KCl}) = 3 \text{ mol/L}$ , (for Ag/AgCl reference systems)



---

**6.2308.040**

**Idrolyte (250 mL)**

Idrolyte (for 6.0224.100 Electrode and Unitrode at  $T > 80^\circ\text{C}$ )



6.2323.000

### Storage solution

Storage solution for all combined pH glass electrodes with reference electrolyte  $c(\text{KCl}) = 3 \text{ mol/L}$



6.2325.000

### pHit kit

Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



6.2325.100

### Cleaning solution 3 x 50 mL

Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode.

The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.



This cleaning solution is also part of the pHit kit.