



# Cu

6.0502.140

ISE

- $\text{Cu}^{2+}$   $10^{-8}$  0.1 mol/L
- 1 mm
- CuEDTA

EP /

## Scope of delivery 6.0502.140

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS


**6.2802.000**

**Polishing set for solid-state electrodes**

Polishing set with 1 polishing cloth and about 2 g aluminum oxide powder (grain size 0.3 µm)



## Optional accessories

Order no.	Description	
6.0750.100	<b>LL ISE reference electrode</b>  Silver / silver chloride reference electrode with double junction system. This reference electrode is well suited for: <ul style="list-style-type: none"><li>• automated applications</li><li>• ion measurements</li><li>• surfactant titrations</li></ul> The ground-joint diaphragm, which is insensitive to contamination, offers a constant and reproducible electrolyte outflow. Additionally, the reference electrolyte is gelified for even better signal stability. The sensor is delivered with $c(\text{KCl}) = 3 \text{ mol/L}$ as bridge electrolyte, which can be freely selected and exchanged as needed.	
6.2104.020	<b>Electrode cable / 1 m / F</b>  For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).	