



Ag/S

6.0502.180

ISE

- $\text{Ag}^+ 10^{-7} \text{ } 1 \text{ mol/L}$
- $\text{S}^{2-} 10^{-7} \text{ } 1 \text{ mol/L}$
- 1 mm
-

EP /

Scope of delivery 6.0502.180

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.2802.000

Polishing set for solid-state electrodes

Polishing set with 1 polishing cloth and about 2 g aluminum oxide powder (grain size 0.3 µm)



Optional accessories

Order no.	Description
6.0750.100	<p>LL ISE reference electrode</p> <p>Silver / silver chloride reference electrode with double junction system. This reference electrode is well suited for:</p> <ul style="list-style-type: none">• automated applications• ion measurements• surfactant titrations <p>The ground-joint diaphragm, which is insensitive to contamination, offers a constant and reproducible electrolyte outflow. Additionally, the reference electrolyte is gelified for even better signal stability. The sensor is delivered with $c(\text{KCl}) = 3 \text{ mol/L}$ as bridge electrolyte, which can be freely selected and exchanged as needed.</p>
6.2104.020	<p>Electrode cable / 1 m / F</p> <p>For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).</p>

