



5 c= 0.7 cm-1 Pt1000

6.0915.100

5 c = 0.7 cm⁻¹ () Pt1000 1.2 m 856
5 µS/cm 20 mS/cm

-
-
-

Scope of delivery 6.0915.100

Qt.	Order no.	Description
-----	-----------	-------------

1 PCS

6.1236.050

Sleeve with SGJ 14/12 mm

Sleeve with SGJ 14/12 mm, polyethylene.



1 PCS

6.2008.040

Storage vessel

Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.



Optional accessories

Order no.	Description
-----------	-------------

6.2301.060 **KCl ion standard 250 mL**

KCl ion standard ($c(\text{ion}) = 0.1000 \pm 0.0005 \text{ mol/L}$), can also be used as standard solution for conductometry 12.87 mS/cm (25°C).



6.2324.010 **Conductivity standard 100 $\mu\text{S/cm}$ 250 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.



6.2324.110 **Conductivity standard 100 $\mu\text{S/cm}$, 5 x 30 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = $0.1/\text{cm}$.



6.2763.000 **Flow-through cell for the 6.0915.100/6.0920.100**

Flow-through cell for the 6.0915.100/6.0920.100 5-ring conductivity measuring cell

